

Certificate of Analysis

ITS Supplement

Catalog No. ITSS-10201-10 Lot Number: T130927G001

Preparation Date: 2013-09-27 Size: 10 ml

Sterility

Bacterial and Fungal Contamination: Samples are inoculated and cultured on blood agar plate, thioglycolate broth, tryptocase soy broth and sabouraud dextrose agar. Specification: No growth must be observed.

Mycoplasma: Samples are tested for mycoplasma contamination using direct culture. Specification: Results must be negative.

Endotoxin: Samples are tested for endotoxin contamination with LAL test. Specification: Results must show a concentration of ≤ 25 EU/ml.

Supporting Capability

Tested for ability to promote proliferation of HeLa cells in serum-free medium.

Results:

All specifications have been met.

Jane Chen

Jane Chen QA Manager Oct 30, 2013